

BRACKET PRO SERIE® 71 USER GUIDE



COPYRIGHT

Copyright 2024 Technimount E.M.S. Holding Inc. (Technimount EMS), an affiliated company of Technologies CGC Inc. All Rights Reserved. No part of this publication may be reproduced, transmitted, transcribed, stored in a retrieval system, or translated into any language in any form or by any means without the written permission of Technimount or its affiliate companies.

Trademark Attributions:

Bracket Pro Serie[®] and Technimount EMS[®] are registered trademarks of Technologies CGC Inc. IntelliVue[™] X3 Monitor is a registered trademark of Philips.

In order to ensure continuous improvements to our products and services, we invite you to communicate your comments to our Customer Service Department at customerservice@technimount.com.

LEGAL NOTICE

Technologies CGC Inc. or Technimount E.M.S. Holding Inc. or its affiliates are not expressly associated with the marketing, distribution or sale of Philips products.

OFFICE ADDRESS

TECHNIMOUNT E.M.S. HOLDING INC. 3770 Jean Marchand Street, Suite 100-C Quebec (QC) G2C 1Y6 Canada


info@technimount.com www.technimount.com T + 1 581.700.6613 TF + 1 888.639.2758

NOTE:

For any issues with your Technimount product, its components, or for any technical questions during the installation or maintenance processes, please send an e-mail to techsupport@technimount.com.

Please have the serial number of your Technimount product available (as shown in the figure below) when calling Technimount Technical Support. Include the serial number in all written communications.

SERIAL NUMBER LOCATION



THE SERIAL NUMBER IS WRITTEN UNDER THE BRACKET AS SHOWN IN THIS IMAGE

TABLE OF CONTENTS

COPYRIGHT	2
LEGAL NOTICE	2
CONTACT INFORMATION	3
Symbols and Definitions	6
Warning/Caution/Note	7
Product Illustration	
BRACKET PRO SERIE 71	
FRONT VIEW BACK VIEW	
INTRODUCTION	-
Purpose of the Document	
Product Description	
Intended Use of the Product	10
TECHNICAL SPECIFICATIONS	
SUMMARY OF SAFETY PRECAUTIONS	12
Warning/Caution/Note	12
NOTICE	14
INSTALLATION GUIDE	15
Removing the Packaging for Installation	15
Tools Required for Installation	15
List of Material	15
Installation of the Dock, monitor and Microstream into the Bracket	
DOCK	
Monitor Microstream and Monitor	
OPERATION GUIDE	-
Removing the Philips IntelliVue X3 Monitor and extension from the Bracket	
Monitor	
Monitor AND microstream	
MAINTENANCE GUIDE	
Cleaning	
CLEANING PROCESS	
Removal of Iodine Compounds	
Preventive Maintenance	23
Inspection Process and Schedule	
Maintenance Program	24
Inspection and Maintenance Record	-
Training Record	
Replacement Parts	
WARRANTY	-
Warranty Policy	
LIMITED RESPONSIBILITY AND WARRANTY	
International Warranty Clause	

TABLE OF CONTENTS

RETURN POLICY	
Return Policy	
Prior to 30 Days	29
Prior 45 Days	29
Prior 60 Days	29
Return Authorization	29
Damaged Merchandise	
Claim Process	
Required Information	
Findings and Conclusion	
Questions about our Policy	

SYMBOLS AND DEFINITIONS

<u>SYMBOL</u>	DESCRIPTION
	Warning and Caution, special attention is required. Consult accompanying documents
	Safe working load symbol and Load balance symbol

WARNING/CAUTION/NOTE


The word warning, caution, or note carry special meaning and should be carefully reviewed.

<u>SYMBOL</u>	DESCRIPTION
	WARNING Alerts the reader about a situation which, if not avoided, could result in death or serious injury. It may also describe potential serious adverse reactions and safety hazards.
	CAUTION Alerts the reader of a potentially hazardous situation, which, if not avoided, may result in minor or moderate injury to the user or patient or damage to the equipment or other property. This includes special care necessary for the safe and effective use of the device to avoid damage that may occur as a result of use or misuse.
NOTE	Provides special information about the product.

PRODUCT ILLUSTRATION

BRACKET PRO SERIE 71


Front View



PRODUCT ILLUSTRATION (CONTINUED)

BRACKET PRO SERIE 71

Back View



PURPOSE OF THE DOCUMENT

This user guide is designed to assist you with the installation, operation and maintenance of the Technimount Bracket Pro Serie 71 for the Philips IntelliVue X3 Monitor.

PRODUCT DESCRIPTION

The Bracket Pro Serie 71 is a mounting system for the Philips Intellivue X3 monitor with or without Microstream Extension. It can be used in different applications such as on a surface, floor, cot/stretcher and wall. It is designed to securely support and transport the Philips Intellivue X3 monitor in pre-hospital environments and it can fit into any ground and air ambulance configurations. When used in emergency vehicles with a cot/ stretcher, the Bracket Pro Serie 71 can be positioned on Technimount's Micro Base, on a Safety MD-Transporter or on vertical mounting systems. These solutions provide a higher flexibility for different EMS/CCT protocols and applications. This mounting system is compatible with all of Technimount's products.

INTENDED USE OF THE PRODUCT

The Bracket Pro Serie 71 is a mounting solution to securely transport the IntelliVue X3 Monitor from Philips during inter-hospital transfers. It is designed to be installed on the equipment as described in this user guide. The Bracket Pro Serie 71 is for crew who has received the appropriate training necessary to operate the equipment in the field according to its intended use as outlined in this user guide. Please read this user guide thoroughly before using this equipment. To ensure the safe operation of this equipment, it is highly recommended to train staff prior to use.

TECHNICAL SPECIFICATIONS

Product Name	Bracket Pro Serie [®] 71	
Product Description	Dual disc bracket for Philips of IntelliVue™ X3 for Emergency and Critical Car Transport	
Compatibility of Medical Device	Philips® IntelliVue™ X3 Monitor, with or without IntelliVue™ Microstream CO₂ Extension	
Environment of Usage	EMS/CCT (Ground) – Hospital/Clinic	
Product Section Material	 Main assembly: Aluminum 6061-T6 and acetal Anti-rotation micro disc: Aluminum 6061-T6 Transport handle: Molded polyurethane 	
Product Finish	Main assembly: Black anodized finish/Anti-rotation micro disc: Silver anodized finish	
Environmental Regulations	RoHS 3 compliant	
Temperature of Usage	-31° F to 113° F (- 35° C to 45° C)	
Certification	Tested in compliance with SAE J3043	
Installation	Device retained on main assembly with red quick release mechanism	
Attachment Options	Anti-rotation micro disc	
Transportation System Compatibility	 Compatible with Technimount's: Safety MD-Transporter[®] (vertical and horizontal) Vertical mounting system Micro base Wall system 	
Features	 Access to all medical device monitors, connectors and accessories Molded polyurethane transport handle Red quick release mechanism for easy installation and removal of medical device Anti-rotation micro disc for installation on vertical and wall mounting systems Adjustment plate for installation of monitor with or without IntelliVue Microstream CO₂ Extension Bottom micro disc allows for 360° rotation when installed on surfaces 	
Physical Overall Dimensions (L x D x H)	8.59 in. x 8.75 in x 7.55 in. (20.32 cm x 20.32 cm x 17.78 cm)	
Weight	4.27 lb (1.94 kg)	
Part Number	330-11-PHX3	
	Contact Technimount EMS for more information or for other options a	

* Product and medical equipment manufacturers names are trademarks[™] or registered trademarks[®] of their respective holders.

WARNING/CAUTION/NOTE

Carefully read and strictly follow all Warnings and Cautions listed in this user guide. Servicing and maintenance should only be done by qualified personnel.

🔨 warning

- ✓ Always hold the bracket firmly when you manipulate it.
- ✓ Operate the bracket only as described in this user guide.
- ✓ Do not modify the bracket, or any components of Technimount systems. Modifying the product can cause unpredictable operation resulting in injury to the patient or operator.
- ✓ It is the responsibility of the operator of the bracket or any of its options, to ensure equipment being used with the Technimount products meets the installation specifications. Injury may result if non-compatible products are used with Technimount System products.
- ✓ Do not attempt to operate, install or remove the equipment while mobile. Ensure that installation and removal of equipment is done while flight or ground emergency service is immobilized.
- Verify that the bracket is inserted correctly into a mounting system and locked. If equipment is improperly installed and locked, it can cause the equipment to fall and cause injury to the patient or operator and/or damage to the medical equipment.
- Never use the bracket if not properly installed and locked on a base, a cot/stretcher, or on a wall mount or floor mount. If not installed properly, the medical device could fall during transport and may cause injury to the patient or operator and/or damage to the product.
- ✓ Do not allow untrained staff to assist in the operation of the mounting system. Untrained technicians/staff can cause injury to the patient or themselves.
- ✓ Do not place additional equipment or items onto the bracket or mounts other than approved Technimount or components, as this could damage the product and may result in injury to the patient or operator.
- ✓ Operators must be able to able to safely lift the total weight of the bracket system and medical device.
- ✓ Do not attach restraints to the bracket frame. It may result in damage to the System and/or result in injury to the patient or operator.
- ✓ Always reposition the equipment with care, avoid any improper pushing or violent manipulation, as it may result in unbalanced equipment and may cause injury to the patient or operator and/or damage to the product.
- ✓ Improper maintenance can cause injury to patient or operator or damage to the product. Maintain the bracket and mounting systems as described in this user guide. Use only Technimount approved parts and maintenance procedures. Using unapproved parts and procedures could cause unpredictable operation and/or injury and will void the product warranty.
- ✓ Failure to properly clean or dispose of corrosive products will increase the risk of premature damage and may cause injury to the patient or operator.
- ✓ Any installation on a surface, wall or on the structure of the vehicle has to be validated or inspected in order to ensure it maintains the integrity of the vehicle.

WARNING/CAUTION/NOTE (CONTINUED)

- ✓ Only certified technical personnel familiar with the bracket installation and use should manipulate it. Always refer to Technimount's user guides for all instructions for installing and using any Technimount Mounting System.
- ✓ Inspect regularly all components on the equipment for any issue and loose screws, bolts and nuts.
- ✓ Never install the mount into or on other similar mounting systems or brackets without written confirmation by Technimount EMS as these different systems may be from other brands or models and may not be compatible with Technimount Systems.

<u>NOTE</u>

- ✓ Some cleaning products are corrosive in nature and may cause damage to the product if used improperly. If such products are used to clean Technimount equipment, the equipment must be rinsed with clean water and thoroughly dried following the cleaning. Failure to properly rinse and dry the Mounting Systems will leave a corrosive residue on the surface, possibly causing premature corrosion of critical components.
- ✓ When ordering, it is important to identify on which mounting system you need to mount the bracket on as there are different systems for different applications (ground and flight).
- ✓ A maintenance program should be established for all Technimount equipment. Preventive maintenance may need to be performed more frequently, based on the level of use of the product. Close attention should be given to safety features (i.e.: screws, quick release mechanisms, and attachment points).
- ✓ Failure to use authorized parts, and lubricants, could cause damage to the equipment and will void the warranty of the product.
- ✓ This user guide should be considered a permanent part of the bracket, and should remain with the product even if the equipment is subsequently sold.
- ✓ Technimount continually seeks advancements in product design and quality. Therefore, while this user guide contains the most current product information available at the time of printing, there may contain minor differences between your mounting system and this user guide.

If you have any questions, or for more information, please contact Technimount Customer Service <u>customerservice@</u> <u>technimount.com</u>.

NOTICE

The Bracket Pro Serie[®] 71 is designed to comply with the essential requirements and other relevant provisions of the SAE J3043. For more information, please contact Technimount Customer Service (<u>customerservice@technimount.</u> <u>com</u>).

TECHNIMOUNT E.M.S. HOLDING INC.

C/O: Regulatory Affairs 3770, Jean Marchand Street, Suite 100-C Quebec (QC) G2C 1Y6 Canada T + 1 581.700.6613 TF + 1 888.639.2758 technimount.com info@technimount.com



The end user and administrator are responsible to validate regulations and standards for safety in their region, in order to comply with applicable safety regulations. Technimount is not responsible to inform the end user and/or administrator of applicable legislation for safety in their area.

REMOVING THE PACKAGING FOR INSTALLATION

- 1. Unpack boxes and check all items are included for installation.
- 2. Ensure that all shipping and packaging materials have been properly removed from the product(s) prior to installation.
- 3. Identify all of the product components and hardware prior to starting the installation.
- 4. The bracket must work properly before it is put into service.
- 5. Refer to the identification of components at the beginning of this user guide if needed.

TOOLS REQUIRED FOR INSTALLATION

Below are the basic tools required for installation. Other tools may also be required. If you have any questions regarding the installation, or need assistance, please contact our Technical Support team at <u>techsupport@technimount.com</u>.

✓ Phillips Screwdriver

<u>NOTE</u>

- ✓ Electrical tools are not recommended, as there is a potential risk of damage to the threads.
- ✓ During the installation process, Technimount recommends the use of medium strength thread-locking adhesive (blue Loctite) to prevent the screws from loosening during normal use. Put a few drops of medium strength threadlocking adhesive (blue Loctite) on each screw before installation.


LIST OF MATERIAL

✓ Bracket Pro Serie 71


INSTALLATION OF THE DOCK, MONITOR AND MICROSTREAM INTO THE BRACKET

DOCK


1. The Bracket Pro Serie 71 is designed to be installed on the Philips IntelliVue X3 Monitor with or without the Microstream CO₂ Extension.



2. The Bracket is shipped fully assembled. To install the Philips Intellivue X3, remove the two (2) screws from the front support plate.




3. Align and insert the IntelliVue X3 Dock into the main assembly.



INSTALLATION OF THE DOCK, MONITOR AND MICROSTREAM INTO THE BRACKET (CONTINUED)

4. Verify that the Dock is inserted properly into the main assembly and ensure that there is no gap between the Dock and the main assembly.



- 5. Align the front support plate with the main assembly.
- 6. Insert the two (2) Black Oxide Stainless Steel Flat Head Phillips Screws 10-32 x 1/2 in.


<u>NOTE</u>

Put a few drops of medium strength thread-locking adhesive (blue Loctite 242) on each screw before installation to prevent the screws from loosening during normal use.

7. Tighten both screws.

🕂 CAUTION


- ✓ Do not use electrical tools to install the screws.
- ✓ Do not overtighten the screws.



INSTALLATION OF THE DOCK, MONITOR AND MICROSTREAM INTO THE BRACKET (CONTINUED)


MONITOR

1. Rotate the adjustable plate to its closed position.



2. Align and insert the Philips IntelliVue X3 Monitor from the right side of the bracket and gently slide into the Dock until you hear a clicking sound that indicates the monitor is locked inside the Dock.

- ✓ Ensure that the monitor is properly inserted and locked by gently pulling on the monitor.
- 3. The Philips IntelliVue X3 Monitor, in its Dock is ready to transport once properly inserted into its bracket and the stretcher mounting system.




INSTALLATION OF THE DOCK, MONITOR AND MICROSTREAM INTO THE BRACKET (CONTINUED)


MICROSTREAM AND MONITOR


1. Rotate the adjustable plate to its opened position.

- Align and insert the IntelliVue Microstream CO₂ Extension from the right side of the bracket and gently slide into the Dock until you hear a clicking sound that indicates the monitor is locked inside the Dock.
- 3. Align and insert the Philips IntelliVue X3 Monitor from the right side of the bracket, and gently slide into the Microstream until you hear a clicking sound that indicates the monitor is locked inside the Microstream.

- ✓ Ensure that the monitor is properly inserted and locked by gently pulling on the monitor.
- 4. The Philips IntelliVue X3 Monitor and the Microstream CO₂ Extension is ready for transport once properly inserted into its bracket and the stretcher mounting system.








REMOVING THE PHILIPS INTELLIVUE X3 MONITOR AND EXTENSION FROM THE BRACKET

MONITOR

1. To remove the Philips IntelliVue X3 Monitor from the bracket, press and hold the Red Quick Release Mechanism.





- 2. While holding the Red Quick Release Mechanism, press on the top Dock button.
- 3. Take the handle of the Philips IntelliVue X3 Monitor and slide it to the right to remove from the bracket.

✓ Always remove the medical device by sliding it to the right. Do not slide it forward.

MONITOR AND MICROSTREAM

4. To remove the Philips IntelliVue X3 Monitor and Microstream from the bracket, press and hold the Red Quick Release Mechanism.






REMOVING THE PHILIPS INTELLIVUE X₃ MONITOR AND EXTENSION FROM THE BRACKET (CONTINUED)

- 5. While holding the Red Quick Release Mechanism, press on the top Dock button.
- 6. Take the handle of the Philips IntelliVue X3 Monitor and slide it to the right to remove from the bracket.

✓ Always remove the medical device by sliding it to the right. Do not slide it forward.



CLEANING

CLEANING PROCESS

ENSURE TO REMOVE MEDICAL DEVICE FROM THE BRACKET IF THOROUGH CLEANING IS NEEDED.

Follow your EMS service's protocol or medical device manufacturer's user guide for the recommended cleaning procedure of your medical devices. Cleaning procedure below does NOT provide recommendations for cleaning of the medical device. The following cleaning guidelines consider the Technimount mounting assemblies and components.

It is the responsibility of each hospital to use the cleaning solution of their choice. All hospitals have policies and procedures describing how they should clean and disinfect their medical equipment. In diluted concentrations, these cleaning agents have no negative effect on our products.

The bracket can be cleaned by using a pressure washer. The bracket may show some signs of oxidation or discoloration from continuous washing, however, degradation of the product's performance or functionality will not be affected due to power washing if cleaning instructions are properly followed.

The material used to manufacture the bracket is high-grade aluminum, it is very durable and resistant to corrosion, however high levels of sodium or a high concentration of sodium or acid may prematurely corrode the product. Always rinse with soft soap and water if exposed.

- ✓ Use of a power washer can accelerate removal of contaminants collected during the use of the product.
- Rinse with clean water.
- ✓ Towel dry all components, and leave to air dry.
- ✓ Avoid over saturation and ensure that the product does not stay wet longer than the cleaner manufacturer's guidelines for proper disinfecting.

When cleaning, always use appropriate personal protection equipment (PPE) based on established protocols (e.g., gloves, eye wear, etc.).

- ✓ Do not steam clean or use ultrasonic cleaners on the system or any of its components.
- \checkmark Do not immerse the metal parts/components in water.
- ✓ Maximum water temperature should not exceed 180°F/82°C.
- ✓ Maximum water pressure should not exceed 500 psi/34,5 BAR. If a pressure washer is being used to clean the unit, the pressure nozzle must be kept a minimum of 24 inches (61 cm) from the unit.

CLEANING (CONTINUED)

CLEANING SOLUTIONS

Phenolic type or quaternary type disinfectants can be used (excluding Virex® TB). Iodophor type disinfectants, Peracetic Acid, Chlorine, and especially Hydrogen Peroxide are not recommended for use because staining and corrosion may result.

Suggested cleaners for the bracket:

- ✓ Quaternary Cleaners (active ingredient ammonium chloride)
- ✓ Water and soft soap
- ✓ A mild 10% bleach solution, a solution containing alcohol or a quaternary ammonium chloride solution. In diluted concentrations, these cleaning agents have no negative effect on our products.

- Some cleaning products are corrosive in nature and may cause damage to the product if used improperly.
- ✓ If the products described are used to clean the equipment, the bracket and components must be rinsed with clean water and thoroughly dried following cleaning. Failure to properly rinse and dry the system will leave a corrosive residue on the surface of the product, possibly causing premature corrosion of critical components.

REMOVAL OF IODINE COMPOUNDS

If rust appears on any surface of the product, we recommend sanding the affected surface to remove the rust and then using the cleaning method presented in this guide once the rust is removed. To clean the stained area, use a solution of 0.13 oz./3.70 ml of Sodium Thiosulfate in 1 pint / .5 litre of warm water to clean the stained area. Clean as soon as possible after staining occurs. If stains are not immediately removed, allow solution to soak or stand on the surface. Thoroughly rinse surfaces with clean water and allow to air dry before returning unit to service.

MARNING

✓ Failure to properly clean contaminated parts or components will increase the risk of exposure to blood borne pathogens and may cause injury to the patient or the operator.

PREVENTIVE MAINTENANCE

A preventive and regular maintenance program should be established for all Technimount equipment. Preventive maintenance may be required more frequently based on the usage level of the product. The bracket requires regular preventative maintenance. Establish and follow a preventive maintenance schedule and keep records of preventive maintenance activities (see maintenance form in this user guide).

- ✓ Close attention should be given to safety features including, but not limited to locking mechanisms and discs.
- Improper maintenance can cause injury or damage to the product. Maintain the product as described in this user guide.

INSPECTION PROCESS AND SCHEDULE

MAINTENANCE PROGRAM

The following inspection routine and schedule is intended as a general guide for preventive maintenance of the bracket. Factors such as weather, environment, geographical location, and individual usage will necessitate different maintenance. If you are unsure as to how to perform these maintenance inspections or at what interval to perform these inspections, please contact your Technical Support Team at techsupport@technimount.com. Check each routine and replace damaged or worn out parts if necessary.

Follow timetable below for maintenance (either in months (M) or hours (H), in accordance with your Service's current maintenance practices and protocols.

<u>ITEM</u>	ROUTINE	<u>1 M</u> OR <u>2 H</u>	<u>3 M</u> OR <u>6 H</u>	<u>6 М</u> <u>OR</u> <u>12</u> <u>Н</u>	<u>12</u> <u>М</u> <u>OR</u> <u>24</u> <u>Н</u>
Bracket	 ✓ Inspect the bracket Inspect the screws Ensure the bracket is working properly and there are no particles, obstacles, or debris ✓ Verify that the Anti-Rotation Micro Disc is clean and works properly ✓ Verify that the Red Quick Release Mechanism works properly ✓ Do a visual inspection of the bracket to ensure there is no damage of chemical attack ✓ Verify that the Bottom Micro Disc is clean and works properly ✓ Verify that the adjustment plate works properly 	X			

INSPECTION AND MAINTENANCE RECORD

PREVENTIVE MAINTENANCE PERFORMED	<u>BY</u>	<u>DATE</u>	<u>TIME</u>

TRAINING RECORD

	DATE		
TRAINEE NAME	<u>BASIC</u> TRAINING	<u>TRAINING</u> <u>UPDATE</u>	<u>TRAINING</u> <u>METHOD</u>

REPLACEMENT PARTS

The parts and accessories listed are all currently available for purchase. Some of the parts identified on the assembly drawing parts in this user guide may not be individually available for purchase. Please call Technimount Customer Service: + 1 888.639.2758 or at <u>customerservice@technimount.com</u> for availability and pricing.

DESCRIPTION	<u>PART NUMBER</u>
Bracket Pro Serie 71	330-11-PHX3
Two (2) Black Oxide Stainless Steel Flat Head Phillips Screws 10-32 x 1/2"	TCM-VF-999-131
Front Support Plate	TCM-092-113-01
Top Locking Assembly	TCM-092-211-01 TCM-092-212-01

WARRANTY POLICY

This statement constitutes TECHNIMOUNT's entire warranty policy with regards to the TECHNIMOUNT Products.

TECHNIMOUNT MAKES NO OTHER WARRANTY OR REPRESENTATION NEITHER EXPRESSED NOR IMPLIED, EXCEPT AS SET FORTH HEREIN. THERE IS NO WARRANTY OF MERCHANTABILITY AND THERE ARE NO WARRANTIES OF FITNESS FOR ANY PARTICULAR PURPOSE. IN NO EVENT SHALL TECHNIMOUNT BE LIABLE HEREUNDER FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES ARISING FROM OR IN ANY MANNER RELATED TO SALES OR USE OF ANY SUCH EQUIPMENT.

TECHNIMOUNT E.M.S. Holding Inc. (TECHNIMOUNT) guarantees to the original "Purchaser" of the "Hardware" with which this "Limited Warranty" is included that the Hardware will be free from "Defects" in workmanship and materials under normal use for a "Warranty Period", one (1) year from the date that the Hardware was first purchased by the Purchaser. During the Warranty Period, the Hardware will be repaired or replaced according to the "Limited Warranty" without charge to the Purchaser for either parts or labor. The parts may be repaired or replaced with either new or refurbished parts. The product may be replaced for a new or refurbished product. For this Limited Warranty "refurbished" means parts and products, which have been returned to factory specifications. If the Hardware is repaired or replaced within the Warranty Period, the greater of the remaining warranty period will apply or three (3) months from the date of repair/replacement. If the Hardware is repaired or replaced after the Warranty Period has expired, the Warranty Period for the repair or replacement will expire three (3) months after the date of repair or replacement.

LIMITED RESPONSIBILITY AND WARRANTY

The Limited Warranty does not apply to normal wear that should result from normal use. It does not apply when the Hardware or any component is opened, disassembled, or repaired by someone not authorized by TECHNIMOUNT and does not cover repair or replacement of any Hardware or part thereof damaged by neglect, misuse, moisture, liquids, exposure to heat, accidents, abuse, and non- compliance with the instructions for installation and use provided with the Hardware. The Limited Warranty does not cover physical damage to the surface of the Hardware, and the decision to repair, replace or refuse the coverage is final and its sole discretion of TECHNIMOUNT, and is made without any compensation or obligation from TECHNIMOUNT.

The hardware, or product defined as a mounting or bracket system for clipping and attaching medical equipment is specifically designed to fill this requirement. Any other use will void the warranty and TECHNIMOUNT will not be held liable on any claim if the product is modified or adapted for use.

Technimount products are intended to retain a medical device in place in the case of a single crash impact. A Technimount product must NOT be used again if it was involved in a crash. The product MUST be replaced. If the Purchaser uses a Technimount product following a crash, it is at the Purchaser's own risk and Technimount will not be held liable.

INTERNATIONAL WARRANTY CLAUSE

This warranty reflects Canadian domestic policy. Warranty outside Canada may vary by country. Please contact TECHNIMOUNT Customer Service for more information.

RETURN POLICY

Mounting systems and bracket systems for portable medical devices may be returned up to 60 days of receipt if:

- The product is not what was originally ordered
- The product does not meet specifications with TECHNIMOUNT's technical sheets according to the customer
- The product is not compatible with the equipment on which it was to be installed

For any manufacturing defect, refer to the conditions of the Warranty Policy.

PRIOR TO 30 DAYS

- Product must be undamaged and in its original packaging
- The product return request must be provided in writing and it must be approved by TECHNIMOUNT prior to returning the product
- Returns will not be approved on a modified or damaged item
- Charges may apply if damaged or missing item
- Customer is responsible for a 10% restocking fee

PRIOR 45 DAYS

- Product must be unused, undamaged and in its original packaging
- The product return request must be provided in writing and it must be approved by TECHNIMOUNT prior to returning the product
- Returns will not be approved on modified or damaged item
- Charges may apply if damaged or missing items
- Customer is responsible for a 25% restocking fee

PRIOR 60 DAYS

- Product must be unused, undamaged and in its original packaging
- The product return request must be provided in writing and it must be approved by TECHNIMOUNT prior to returning the product
- Returns will not be approved on a modified or damaged item
- Charges may apply if damaged or missing items
- Customer is responsible for a 30% restocking fee

RETURN AUTHORIZATION

TECHNIMOUNT's customer service department is responsible for approving any merchandise return and will provide a Return Merchandise Authorization (RMA) number to be printed on any returned merchandise. TECHNIMOUNT reserves the right to charge shipping and restocking fees on returned items. SPECIAL, MODIFIED, OR DISCONTINUED ITEMS NOT SUBJECT TO RETURN.

DAMAGED MERCHANDISE

ICC Regulations require that claims for damaged merchandise must be made with the carrier within fifteen (15) days of receiving merchandise. DO NOT ACCEPT DAMAGED SHIPMENTS UNLESS SUCH DAMAGE IS NOTED ON THE DELIVERY RECEIPT AT THE TIME OF RECEIPT. Upon prompt notification, TECHNIMOUNT will file a freight claim with the appropriate carrier for damages incurred. Claim will be limited in amount to the actual replacement cost. In the event that this information is not received by TECHNIMOUNT within the fifteen (15) day period following the delivery of the merchandise, or the damage was not noted on the delivery receipt at the time of receipt, the customer will be responsible for payment of the original invoice in full.

Note: Claims for any short or broken shipment must be made within thirty (30) days of invoice.

CLAIM PROCESS

Please follow the claim procedure below to return a defective product: Fill in the RMA form online at https://www.technimount.com/service-support/service-support-overview/

REQUIRED INFORMATION

- Serial number
- Invoice number
- Identification of the issue
- Upload the pictures of the issue
- Submit the form

A confirmation of the receipt of the claim will be sent to you within 2 business days. If you have not received a confirmation e-mail by then, please contact <u>techsupport@technimount.com</u>.

If the RMA request is approved, TECHNIMOUNT will provide you with an RMA number for your claim. We will also include instructions for the replacement of your product. No product return will be accepted without the RMA number labeled on the outside of the box being returned.

FINDINGS AND CONCLUSION

- If following an investigation, we find that our product is defective, TECHNIMOUNT will take corrective actions and close the claim
- If following an investigation, we find that our product is NOT defective (misuse or abuse of the product) the product will not be covered by the warranty. Details of our findings and conclusions will be provided to you

QUESTIONS ABOUT OUR POLICY

If you have any questions regarding our return policy, please contact our Customer Service department at <u>customerservice@technimount.com</u>.



SAFETY AND FLEXIBILITY WHERE IT MATTERS MOST

TECHNIMOUNT E.M.S. HOLDING INC.

3770 Jean-Marchand Street, Suite 100-C Quebec (QC) G2C 1Y6 Canada T + 1 581.700.6613 TF + 1 888.639.2758 technimount.com info@technimount.com

© 2024 Technimount E.M.S. Holding Inc. - All Rights Reserved Printed in Canada



UG-BP71-202407EN-04